



## CALIBRATION CERTIFICATE

**CUSTOMER:** 4 Star Ventures **DATE:** May 9, 2014  
**CERTIFICATE#** 29740-30161 **SERIAL#** 00331  
**MANUFACTURER:** Schwabs **P.O.#**  
**MODEL:** 6" Joint Torque Assembly **CYLINDER:** 4.409 In<sup>2</sup>  
**TEST GAUGE#** 004 **CYL. S/N#** 00800A  
**RATIO:** 25" Handle

**CALIBRATION PRESSURE**

psi  
327  
653  
980  
1306

**GAUGE DIAL READING**

ft.lbs  
3000  
6000  
9000  
12000

**ACTUAL DIAL READING**

ft.lbs  
3000  
6000  
9000  
12000

**TECHNICIAN:** Tam Pham

**COMMENTS:**

WE HERBY CERTIFY THAT THE ABOVE INSTRUMENT WAS CALIBRATED ACCORDING TO MANUFACTURER'S SPECIFICATIONS WHICH ARE +/- 2% OF FULL SCALE. THE CALIBRATION STANDARD IS TRACEABLE TO THE CANADIAN NATIONAL BUREAU OF STANDARDS AND COMPLIES WITH CANADIAN STANDARDS ASSOCIATION CAN3-299-3-85 AND A.N.S.I. B40-1.

**Kane Instrumentation Ltd.**  
2109 - 84 Avenue  
Edmonton, Alberta T6P 1K1  
(780) 416-5263 (780) 416-8309 fax